



What is the difference between weak and strong cation exchange for Ion Exchange columns - FAQ

An HPLC stationary phase is considered a weak cation exchanger when it has a functional group such as a carboxylic acid which proton does not come off (ionize) until around pH 4 or so.

Columns with sulfonic acid groups for example, are considered strong cation exchangers because they are essentially un-protonated (ionized) at any reasonable mobile phase pH for HPLC.



Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com