



Will phosphates precipitate in an HPLC column - FAQ

Phosphoric acid will not precipitate in the columns. Sodium phosphate (which is not recommended) but which is often used as a buffer, will precipitate in the columns and should be avoided as a storage solvent for all HPLC columns.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com