



Why would I want to use an HPLC column that is not end capped - FAQ

End-capping the stationary phase is a great idea for neutral and non-polar compounds that are hydrophobic. This is common with traditional Reversed Phase stationary phases.

Polar compounds do not retain well enough on these end-capped columns but might retain and separate on non end-capped columns due to the increased polar nature of the stationary phase. However, there are better options.

While the Cogent TYPE-C™ columns do not require to be end-capped, they can retain both polar and non polar compounds in Reversed Phase or Aqueous Normal Phase ANP.

[Click here to learn about TYPE-C HPLC columns](#)



Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com