

Are the MicroSolv mobile phase bottles Type 1, Class A glass - FAQ

Yes, they are.

Both the safety-coated and non-coated MicroSolv mobile phase bottles are crafted from high-quality borosilicate 3.3 glass, a type of Type 1, Class A borosilicate glass. This material offers excellent chemical resistance, making these bottles suitable for a wide range of laboratory applications.

- Click <u>HERE</u> for ordering information for safety coated bottles.
- Click **HERE** for ordering information for non coated bottles.

Printed from the Chrom Resource Center
Copyright 2025, All Rights Apply
MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900
Fax: (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com