


# Ethanolamine Analyzed with LCMS - AppNote

# Retention & Analysis of Hydrophilic, Ethanolamine from Plasma

In this Application Note, the Chromatograms show Ethanolamine sufficiently retained using either the Gradient or Isocratic Methods. The advantages of using these Methods are fast equilibration between runs when Gradient analysis is required and due to the high organic content of the Mobile Phase, the Ionization efficiency of MS Detector produces increased sensitivity when compared to Methods with a high Water content Mobile Phase.



## **Method Conditions**

Column: Cogent Diamond Hydride™, 4µm, 100Å

**Catalog No.**: <u>70000-15P-2</u> **Dimensions**: 2.1 x 150mm

Mobile Phase: Gradient Figure A

A: 50:50 DI Water / Isopropanol with 0.1% Formic Acid (v/v) B: 97:3 Acetonitrile / DI Water with 0.1% Formic Acid (v/v)

Isocratic Figure B

50:50 Solvent A / Solvent B

### Gradient:

Time (minutes)	%B
0	80
4	40
6	40
7	80

Post Time: 3 minutes Injection vol.: 1µL

Flow rate: 0.4mL / minute

**Detection**: ESI – POS - Agilent 6210 MSD TOF Mass Spectrometer

to: 0.9 minutes

Note: Ethanolamine is bi-functional, containing both a primary amine and a primary alcohol. Colorless, viscous and with an odor reminiscent of ammonia, its derivatives are widespread in nature, including: modulating several animal and plant physiological processes such as seed germination, plant—pathogen interactions, chloroplast development and flowering. It is also a precursor and when combined with arachidonic acid the endocannabinoid Anandamide is formed.

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com



#### **Attachment**

No 210 Ethanolamine Analyzed with LCMS pdf 0.3 Mb Download File

Printed from the Chrom Resource Center Copyright 2025, All Rights Apply **MicroSolv Technology Corporation** 9158 Industrial Blvd. NE, Leland, NC 28451