



Separation of Polar Solutes with HPLC - AppNote

Niacin, Riboflavin, Folic Acid, Pyridoxine, Metformin, Thiamine

The Cogent Diol Column is a good addition to the TYPE-C™ Silica line of HPLC stationary phases. Here, a variety of common polar analytes are well-retained and separated.





Peaks:

1. Ascorbic acid
2. Niacin
3. Riboflavin
4. Folic acid
5. Pyridoxine
6. Metformin
7. Thiamine

Method Conditions

Column: Cogent Diol™, 4μm, 100Å

Catalog No.: [40060-15P-3](#)

Dimensions: 3.0 x 150mm

Mobile Phase:

A: DI Water / 0.1% Formic Acid (v/v)

B: Acetonitrile / 0.1% Formic Acid (v/v)

Gradient:

Time (minutes)	%B
0	95
3	95
10	40
12	40
13	95

Post Time: 5 minutes

Flow rate: 0.7 mL/minute

Detection: UV @ 254 nm

Injection vol.: 1μL

Sample Preparation: Mixture of reference standards in diluent of 50 / 50 Solvent A / Solvent B.

t₀: 0.7 minutes

Note: B and C Vitamins are hydrophilic and therefore may be difficult to retain in Reversed Phase methods. Metformin is a highly polar compound used for treatment of type 2 diabetes.



Attachment

No 291 Separation of Polar Test Solutes pdf 0.3 Mb [Download File](#)

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com