



## How are MicroSolv brand vials packaged to ensure no contamination - FAQ

All MicroSolv brand vials that carry a CleanPack™ label on the front side of the polypropylene box have been packed in a certified clean room, class 10,000 after having passed the Lehrer oven at approximately 600°C.

You consumer can rely on clean, uncontaminated vials for a correct analysis.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)