



## Will I see any differences in my HPLC chromatography if I change to AQ syringe filters - FAQ

**If you use proper filtering techniques and a proper filter you should not see any differences in your chromatography, CE or UV testing.** Syringe filters only protect your equipment and columns and never participate in separations or SPE.

Click [HERE](#) for syringe filters ordering information and pictures.

### Attachment

**MicroSolv filters equivalency study pdf** [Download File](#)

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)