



Your filter sparging stones with stems are listed as 1/8 or 1/16 ID shouldn't this be OD - FAQ

In the case of certain types of filter sparging stones where the tubing is inserted into the stone, the ID of the stems fits the OD of the low pressure tubing snugly.

The following schematic should clarify the above sentence. The ID of the sparging stone stem and the OD of the low pressure tubing are the same (**red** lines). The OD of the sparging stone stem (**green**) is larger than the ID of the low pressure tubing (**blue**).



In another type of sparging stone, the OD of the sparging stone stem and the ID of the low pressure tubing are the same (**blue** lines). The ID of the sparging stone stem (**orange**) is smaller than the OD of the low pressure tubing (**red**).



Click [HERE](#) for filter sparging stone ordering information and pictures.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com