



What is the maximum temperature for Cogent Diamond Hydride HPLC columns - FAQ

The maximum operating temperature for the Cogent Diamond Hydride™ HPLC column is 60° C.

However, Higher Temperature Columns are available on request with High Temperature all Stainless Steel Hardware.



[Click HERE for Cogent Diamond Hydride Ordering Information](#)

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com