



My method uses an standard type-B silica based HPLC column is there a Cogent TYPE-C column equivalent substitute - FAQ

While many columns use the same nomenclature for their stationary phases such as C18, phenyl, C8 etc, every brand of column has unique properties and in any given method, columns can perform equally and then in another method perform differently.

It is the belief of MicroSolv Technology Corporation that our Cogent TYPE-C™ columns are unique in the way they perform and therefore, we do not offer our products as equivalents. That is not to say that our columns will not perform equally or better in any method that can be optimized for our columns that are properly chosen for the sample.



Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com