

Can I use Silica C columns in a Normal Phase gradient method - FAQ

Yes!

Although this is not easy with ordinary silica columns due to solvent de-mixing & water uptake, the Cogent Silica-C™ column is not limited by this and is easy to use even with fast gradients.



Printed from the Chrom Resource Center Copyright 2025, All Rights Apply **MicroSolv Technology Corporation** 9158 Industrial Blvd. NE, Leland, NC 28451 Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com