

Can the HSQ Kit be used to measure the LOQ and LOD differences between instruments - FAQ

Yes, this is another really great “extra feature” of the HSQ Kit™.

By determining the, signal to noise ratio, the extra column volumes and the dwell volume of each instrument using the HSQ Kit™ and the software provided, you can determine the differences in LOQ (limits of quantitation) and LOD (limits of detection) between instruments. This can be very helpful in troubleshooting lack of precision in methods that transfer from one instrument to the other.

Click [HERE](#) for HSQ Kit ordering information and pictures.

Printed from the Chrom Resource Center
Copyright 2025, All Rights Apply
MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900
Fax: (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com