



Do different retention times for peaks in the Resolution Test Mixture on different systems impact the integrity of my data - FAQ

No impact at all.

There will certainly be mobile phase preparation differences, aging of columns, etc. The HSQ Kit™ instructions call for adjustment of the mobile phase to meet system suitability, which for example in our C of A, calls to adjust the mobile phase to achieve a 4-5 minute retention time window. Keep in mind that the C o A merely confirms that the column was tested, and met specs when it left our facility.

Click [HERE](#) for PQ and HSQ Kit ordering information and pictures.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com