

## Do Partisil HPLC columns need an extra end fitting kit - FAQ

**No, Partisil™ HPLC columns are a complete column and no extra end fitting kit is required.**

This column has the standard compression end fittings, WCS.

**It is the Partisphere™ which is a cartridge type HPLC column and needs an end fitting kit.**

It has a unique void sealing hardware named WVS. All WVS cartridge columns are shipped *without* end fittings and an require end fitting kit. The idea of a cartridge system is to have one or two sets of end fittings that are removable and can be used interchangeably on the cartridges as you install them on the instrument.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)