



Analyzing metformin in plasma samples how do I obtain reproducible results and peak shapes - FAQ

Suggestion: Try substituting formic acid in DI water and adding ammonium acetate (10mM) and acetic acid (1%) to the aqueous phase. This should improve the reproducibility of peak areas and retention times.

In addition, extensive reconditioning of the analytical column with pure acetonitrile in between runs is essential for good reproducibility when working with plasma samples.



Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com