PRODUCT SHEET

MICROSOLV
TECHNOLOGY CORPORATION
HPLC Columns and Chromatography Accessories

Catalog Number

47450-05-005

Adapter, for use with Low to Medium Pressure HPLC. 10-32 female to 1/4"-28 female connection, 0.25 mm bore size. PEEK. ARE-Applied Research brand. 5/PK.

Additional Info

This End Fitting is made for 2.1mm ID PEEK Columns with standard HPLC column 10-32 (tubing inlet) screw threads and 1/4"-28 "column tube" screw threads. The through hole is 0.25mm.

Designed for use only with our All-PEEK HPLC Column Hardware and not recommended for use with stainless steel columns.

Machined from PEEK which ensures a biocompatibility from the beginning of the packing and column use.

MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451
tel. (732) 380-8900, fax (910) 769-9435
Email: info@mtc-usa.com
Website: www.mtc-usa.com