PRODUCT SHEET

MICROSOLV
TECHNOLOGY CORPORATION
HPLC Columns and Chromatography Accessories

Catalog Number

49911-01

Ferrules, Replacement, for Di-Ad, Direct Adaptive HPLC Column Connectors. PEEK. MICROSOLV Brand. 10/PK.

Additional Info

Replacement ferrules for Direct Adaptive (Di Ad™) HPLC Column Connectors are designed to maintain reliable sealing performance and consistent low dead volume fluidic connections in analytical HPLC systems.

The 49911 01 replacement ferrules are manufactured from all PEEK material and are intended for use with MICROSOLV Di Ad™ HPLC Column Connectors after extensive routine use and repeated column changes.

After several hundred connections to HPLC columns or instrument ports, replacing the ferrule is recommended to help maintain dependable sealing performance and reduce the possibility of leaks during operation.

These replacement ferrules are easy to remove and install, allowing convenient maintenance of existing Di Ad™ connector assemblies without replacing the entire fitting.

Key Benefits

MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451
tel. (732) 380-8900, fax (910) 769-9435
Email: info@mtc-usa.com
Website: www.mtc-usa.com