PRODUCT SHEET

MICROSOLV TECHNOLOGY CORPORATION
HPLC Columns and Chromatography Accessories

Catalog Number

58022-N13-K

Syringe Filters, 13mm, Nylon, 0.22um Pore Size. Purple Polypropylene, AQ Brand. 1000/CS.

Additional Info

Advanced Quality Nylon-66 Syringe Filters, AQ™ brand,for Pharmaceutical Grade HPLC Dissolution Testing. Manufactured with a unique 0.22µm nylon-66 membrane in a 13mm polypropylene housing, these syringe filters are made for laboratories that require consistent, high performance results with ease of use especially during high volume work. Excellent choice when laboratories cannot afford poor chromatographic results.

These filters are compatible with aqueous or organic HPLC solvents and have been designed to be extremely low adsorbing even for charged molecules and extractable free even in the first 1ml of filtration.

Validated for many regulatory requirements, they are designed, tested and QC certified. They comply with US FDA CFR Title 21, Part 210.3(b)(6) for minimum Fiber Release and with USP <788> test for particulate matter in injectables as well as USP for "Sterile Water for Injection" and therefore do not require any initial flushing or discarding of the first 200µl of filtrate saving time and avoiding investigation of failed methods. The polypropylene plastic housing material complies with the requirements of US FDA CFR Title 21 Part 177.1520 are heat sealed ...

MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451
tel. (732) 380-8900, fax (910) 769-9435
Email: info@mtc-usa.com
Website: www.mtc-usa.com